



# User Manual

## OCTANE SUB4



WHEELCHAIR  
DIRECTIONS FOR USE

# Thank you for choosing RGK,

Welcome to the RGK family and we hope the product and service you have received has exceeded your expectations. Please take the time to read the instructions and familiarize yourself with your new wheelchair to ensure you benefit from the maximum safety, performance and longevity.

We want you to be fully satisfied with our products and services and we are constantly working to develop our products further. For this reason, changes can occur within our range of products with regard to form, technology and equipment. Consequently, no claims can be construed from the data or pictures contained in this user manual.

## **WARNING:**

Do not operate this equipment without first reading and understanding this user manual. If you are unable to understand the warnings and instructions, contact your Sunrise Medical authorized dealer or a healthcare professional (doctor / therapist) before attempting to use this equipment - otherwise injury or equipment damage may result.

## Set Up Guide:

1. Carefully remove product and any loose parts from the box and ensure nothing is left in the packaging.
2. Read contents of this user manual thoroughly and understand the safety risks.
3. Attach the wheels and separate parts.
4. Check everything is as expected and that all parts are secured, functioning and set up correctly, including tire pressure.
5. Adjust the upholstery ensuring it provides the right levels of support to suit your posture and comfort requirements.
6. Adjust the footplate height if required to ensure the correct support for your feet and balance.
7. Once ready and in a safe environment, carefully transfer into your new wheelchair and ensure it functions correctly.
8. Before going outdoors or on rougher terrain, ensure you have fully tested the wheelchair and are confident with the balance point and how the chair functions.
9. If you have questions on the set up of your wheelchair, please contact your local authorized dealer. Happy rolling!

Sunrise Medical (US) LLC  
North American Headquarters  
2842 N. Business Park Avenue  
Fresno, CA, 93727, USA  
(800) 333-4000  
(800) 300-7502  
[www.SunriseMedical.com](http://www.SunriseMedical.com)

# Contents

4	Daily Manual Propelled Wheelchair
5	Safety Summary
8	Safety Inspection Checklist
9	Troubleshooting Guide Mechanical
10	Procedure 1 - Upholstery
11	Procedure 2 - Rear Wheels
12	Procedure 3 - Tires and Inner Tubes
12	Procedure 4 - Camber Bar
13	Procedure 5 - Caster Wheels and Forks
13	Procedure 6 - Wheel Locks
14	Procedure 7 - Footrest
14	Procedure 8 - Folding Backrest
15	Procedure 9 - Push Handles
16	Procedure 10 - Anti-tips
17	Procedure 11 - Sideguards
18	Transportation
18	Technical Data
18	Recycling/Disposal
19	Warranty

## Definitions

### NOTE:

General advice or best practice

### ⚠ CAUTION:

Advice to user that potential damage to equipment may occur if the advice is not followed.

### ⚠ WARNING:

Advice to the user of a Potential Risk of injury if the advice is not followed.

### ⚠ DANGER:

Advice to the user of a potential risk of serious injury or death if the advice is not followed.

### NOTE:

The wheelchairs shown and described in this user manual may not correspond in every detail exactly to your own model. However, all instructions are completely relevant, regardless of possible detail differences. The manufacturer reserves the right to alter without notice any weights, measurements or other technical data shown in this manual. All figures, measurements and capacities shown in this manual are approximate and do not constitute specifications.

### NOTE:

Please keep a note of your local authorized dealer's contact details in the space provided. In the event of a breakdown, contact them and try and give relevant details so they can help you quickly.

---

---

---

Fig 2.



- |                          |                                 |
|--------------------------|---------------------------------|
| <b>A</b> - BACKREST      | <b>G</b> - FOOTPLATE COVER      |
| <b>B</b> - WHEEL LOCKS   | <b>H</b> - CAMBER AXLE          |
| <b>C</b> - REAR WHEELS   | <b>I</b> - FRAME                |
| <b>D</b> - CASTER WHEELS | <b>J</b> - UNDER SEAT BRACE BAR |
| <b>E</b> - CASTER FORKS  | <b>K</b> - CUSHION & COVER      |
| <b>F</b> - FOOTPLATE     | <b>L</b> - SIDEGUARD            |

# Safety Summary

## Use

Wheelchairs are exclusively for children and adults who are unable to walk or have limited mobility, for their own personal use, when self-propelling and being moved by a third party (pushed by an attendant), within the home and outdoors. The maximum weight limit (includes the user and any weight of accessories fitted to the chair) is marked on the serial number label, which is affixed to the camber axle or the under-seat brace bar.

Warranty can only be considered if the product is used under the specified conditions and for the intended purposes. The intended lifetime of the wheelchair is 5 years. Please do not use or fit any third-party components to the wheelchair unless they are officially approved by Sunrise Medical.

## Area of Application

The variety of fitting variants as well as the modular design mean that it can be used by those who cannot walk or have limited mobility e.g. due to:

- Paralysis
- Loss of extremity (leg amputation)
- Extremity defect deformity
- Joint contractures / joint injuries
- Illnesses such as heart and circulation deficiencies, disturbance of equilibrium or cachexia as well as for elderly people who still have the strength in the upper body.

When considering provision, please also note the body size, weight, physical and psychology constitution, the age of the person, living conditions and environment.

Final Selection of the type of wheelchair to be used by an individual rests solely with the user and their health care professional capable of making such a selection.

## WARNING:

CHOKING HAZARD – This mobility aid uses small parts which under certain circumstances may present a choking hazard to young children.

## General safety

The engineering and construction of this wheelchair has been designed to provide maximum safety. International safety standards currently in force have been fulfilled where appropriate. Nevertheless, users may put themselves at risk by improperly using their wheelchairs. For your own safety, the following rules must be strictly observed.

Unprofessional or erroneous changes or adjustments increase the risk of accident. As a wheelchair user, you are also part of the daily traffic on streets and pavements, just like anyone else. We would like to remind you that you are therefore also subject to any and all traffic laws.

Be careful during your first use in your new wheelchair and get comfortable and confident before attempting outdoors or difficult obstacles. It is possible that potholes or uneven ground could cause this wheelchair to tip over, especially when riding uphill or downhill. When riding forwards over a step or up an incline, the body should be leaning forward.

## WARNING:

Travelling downhill at higher speeds in your wheelchair can cause caster flutter. This can lead to instability and potentially result in personal injury.

Before each use, the following should be checked:

- Quick release axles on the rear wheels
- Hook and loop patches on seats and backrests
- Tires and tire pressure
- Wheel locks

Before changing any of the adjustments of the wheelchair, it is important to read the corresponding section of the user manual.

## DANGER:

NEVER exceed the maximum load of 242 lbs (110kg) for Octane Sub4 (187 lbs [85kg] if used with a handbike) for driver plus any items carried on the wheelchair. Please note the weight information for lighter weight options, which are quoted separately. If you exceed the maximum load, this can lead to damage to the chair, or you may fall or tip over, lose control and may lead to serious injury of the user and other people.

# Safety Summary Continued

- If and whenever possible, during a journey in a specially fitted vehicle for disabled people, vehicle occupants should use the seats in the vehicle and the appropriate restraint system. This is the only way to ensure that occupants will have the maximum protection if there is an accident.
- To avoid falls and dangerous situations, you should first practice using your new wheelchair on level ground with good visibility.
- This wheelchair should be used exclusively to convey one person at a time. Any other use does not comply with the intended purpose.
- When getting on or off the wheelchair, do not use the footplates. These should be flipped up beforehand and swung to the outside as far as possible. (If applicable)
- Depending on the diameter and setting of the casters, as well as the center of gravity setting of the wheelchair, the casters may begin to flutter at high speeds. This can lead to the casters being blocked and the wheelchair may tip over.
- In particular, do not travel on an incline without wheel locks, travel at a reduced speed. Riding sideways on to a slope or incline, increases the possibility of the wheelchair tipping over sideways.
- Explore the effects of changing the center of gravity on the behavior of the wheelchair, for example on inclines, slopes, all gradients or when overcoming obstacles. Do this with the secure aid of a helper.
- We recommend that novice users use anti-tip tubes.

## Weight Training

Sunrise Medical DOES NOT recommend the use of its wheelchairs as a weight training apparatus. This product has NOT been designed or tested as a seat for any kind of weight training. If occupant uses said wheelchair as a weight training apparatus, Sunrise Medical shall NOT be liable for personal injury or equipment damage and the warranty will be voided immediately.

### **⚠ WARNING:**

Any products left in extreme hot or cold climates could result in hot/cold surfaces which could result in personal injury.

Do not expose the product or any of its accessories to any form of excessive heat or naked flame, including cigarettes.

This product has moving parts and therefore could trap the user's fingers. Use extreme caution when moving.

## Stability and Balance

To assure stability and correct operation of your wheelchair, you must at all times maintain balance. Your wheelchair has been designed to remain upright and stable during normal daily activities as long as you do not move beyond the center of gravity. All activities which involve movement in the chair have an effect on the center of gravity.

Make sure the casters are pointing in the rearward position whenever you lean forward. This can be achieved by advancing the chair and then reversing it in a straight line. See Fig 1.

Fig 1.



### **⚠ WARNING:**

The position of the footrest, camber tube and rear seat height, as well as the tautness of the back upholstery are directly related to the chair's stability. Any change to one or any combination of the three may cause the chair to decrease in stability. Use EXTREME caution when using a new seating position for the first time.

### **⚠ WARNING:**

#### **Tire pressure**

DO NOT use your wheelchair unless it has the proper tire pressure (P.S.I.).

DO NOT over inflate the tires. Failure to follow these suggestions may cause the tire to explode and cause bodily harm.

# Safety Summary Continued

- Anti-tip tubes should prevent the chair tipping over backwards unintentionally. Under no circumstances should they take the place of transit wheels, and be used to transport a person in a wheelchair with the rear wheels removed.
- With extreme settings (e.g. rear wheels in the most forward position) and less than perfect posture, the wheelchair may tip over even on a level surface.
- Do not hang heavy items such as shopping bags or backpacks, on the push handles, headrest or backrest of the wheelchair. These can change the tipping point and there is a risk of tipping backwards.
- When reaching for objects which are in front of, to the side or behind the wheelchair, make sure that you do not lean too far out of the wheelchair. If you do it can change the center of gravity and there is a risk of tipping or rolling over.
- Only use your wheelchair properly. For example, avoid travelling against an obstacle without braking (step, curb edge) or dropping down gaps, or using escalators.
- Only attempt stairs with the help of an attendant. There is equipment available to help you, e.g. climbing ramps or lifts, please use them. If there is no such equipment available, then the wheelchair must be tipped and pushed over the steps (2 helpers).
- In general, any anti-tip tubes fitted must be set beforehand, so that they cannot touch the steps, as otherwise this could lead to a serious fall. Afterwards the anti-tip tubes must be set back to their correct position.
- Make sure that the attendant only holds the wheelchair using securely mounted parts (e.g. not on the footrests or the side guards).
- When using the lifting ramp make sure that the anti-tip tubes fitted are positioned outside the danger area.
- Secure your wheelchair on uneven ground or when transferring (e.g. into a car) by using the wheel locks.
- The wheel locks will only work if there is sufficient tire pressure and if the correct setting has been made (see the Chapter on "Wheel locks").
- When it is dark, please wear light clothing or clothing with reflectors, so that you can be seen more easily. Make sure that the reflectors on the side and back of the wheelchair are clearly visible. We would also recommend that you fit an active light.
- Adjustments to your wheelchair, especially to security-relevant components as for example wheel-locks, anti-tip or backrest must be performed by authorized dealers.
- It is recommended to clean and disinfect the entire product, especially the upholstery, frequently.
- Do not tilt the wheelchair without assistance.
- Do not reuse single use fasteners.
- The wheel locks are not intended to brake your wheelchair. They are only there to ensure that your wheelchair does not begin rolling unintentionally. When you stop on uneven ground, you should always use the wheel locks to prevent such rolling. Always apply both wheel locks; otherwise, your wheelchair could tip over.
- The wheel locks have not been designed to be used as brakes for a moving wheelchair.
- Always make use of elevators and ramps. If these are not available, you should request the help of an attendant. They should only grip the wheelchair at securely mounted parts. If anti-tip tubes are fitted these must be folded away. A wheelchair should never be lifted with an occupant; it should only be pushed.
- If the seat and back sling are damaged, you must replace them immediately.
- Do not lift or carry the wheelchair by the back tubes or the push handles.
- Be careful with fire, in particular with burning cigarettes. Seat and back slings can be set alight.
- Always make sure that the quick-release axles on the rear wheels are set properly and are locked in place. If the button on the quick-release axle is not pressed in, the rear wheel cannot be removed.
- In particular when using lightweight metal handrims, fingers will easily become hot when braking from a high speed or on long inclines.
- If the wheelchair is subject to direct sunlight for a long period of time, then parts of the wheelchair (e.g. frame, leg rests, brakes and side guards) may become hot (>106°F [>41°C]).
- To avoid hand injuries do not grab in between the spokes or between the rear wheel and wheel lock when driving the wheelchair.
- Do not use the wheelchair on slopes > 6°.
- The dynamic safe slope is dependent on the chair configuration, the user's abilities and the style of riding. As the user's abilities and style of riding cannot be pre-determined then the max safe slope cannot be determined. Therefore, this must be determined by the user with the assistance of an attendant to prevent tipping. It is strongly recommended that inexperienced users have anti-tips fitted.
- Do not use your wheelchair on muddy or icy ground.
- Do not use your wheelchair where pedestrians are not allowed.
- The wheelchair should be used with caution in heavy rain, snow, slippery, or unsound surfaces.
- Do not use the wheelchair in hazardous environments.

# Safety Inspection Checklist

## NOTE:

Regular cleaning will help to highlight any loose or worn parts and enhance the smooth operation of your wheelchair. To operate properly and safely, your wheelchair must be cared for just like any other vehicle. Routine maintenance will extend the life and efficiency of your wheelchair. Initial adjustments should be made to suit your personal body structure and preference. Thereafter follow these maintenance procedures. Outlined below:

Item	Initially	Inspect/ Adjust Weekly	Inspect/ Adjust Monthly	Inspect/ Adjust Yearly
<b>General</b> <ul style="list-style-type: none"> <li>Wheelchair rolls straight (no excessive drag or pull to one side)</li> <li>Inspect frame for cracks and damage</li> <li>Inspect all fasteners are secured</li> </ul>	✓		✓	✓
<b>Wheel Locks</b> <ul style="list-style-type: none"> <li>Do not interfere with tires when rolling</li> <li>Pivot points free of wear and looseness</li> <li>Wheel locks easy to engage</li> </ul>	✓ ✓ ✓		✓ ✓ ✓	
<b>Seat and Back Upholstery</b> <ul style="list-style-type: none"> <li>Inspect for rips or sagging</li> <li>Inspect fastening flaps to ensure they securely latch</li> </ul>	✓ ✓		✓	✓
<b>Rear Wheels</b> <ul style="list-style-type: none"> <li>Adjustable axle position camber bar is securely tightened</li> <li>Quick/quad release axles lock properly</li> <li>No excessive side movement or binding when lifted and spun</li> </ul>	✓ ✓ ✓	✓ ✓	✓	
<b>Handrims</b> <ul style="list-style-type: none"> <li>Inspect for signs of rough edges or peeling</li> </ul>	✓	✓		
<b>Spokes</b> <ul style="list-style-type: none"> <li>Inspect for bent or broken spokes</li> <li>All spokes uniformly tight</li> </ul>	✓ ✓		✓ ✓	
<b>Front Caster</b> <ul style="list-style-type: none"> <li>Inspect wheel/fork assembly for proper tension by spinning caster; caster should come to a gradual stop</li> <li>Loosen/tighten locknut if wheel wobbles noticeably or binds to a stop</li> <li>Wheel bearings are clean and free of moisture</li> </ul> <p>CAUTION: As with any vehicle, the wheels and tires should be checked periodically for cracks and wear, and should be replaced when necessary</p>	✓ ✓ ✓		✓ ✓ ✓	✓
<b>Tires</b> <ul style="list-style-type: none"> <li>Inspect for flat spots and wear</li> <li>If pneumatic tires check for proper inflation</li> </ul> <p>CAUTION: As with any vehicle, the wheels and tires should be checked periodically for cracks and wear, and should be replaced when necessary</p>	✓ ✓	✓ ✓		
<b>Cleaning</b>	✓		✓	



# Troubleshooting Guide Mechanical


Chair veers right	Chair veers left	Slow turn or performance	Caster flutters	Squeaks and rattles	Looseness in chair	Solutions
✓	✓	✓	✓			Check tires for correct and equal pressure
		✓	✓	✓	✓	Check for loose caster stem nuts Check all fixtures are secure and tight
				✓	✓	Check spokes for tension
✓	✓		✓			Check that both casters contact the ground at the same time

## **WARNING:**

After adjusting, always make sure that parts are properly tightened BEFORE using the wheelchair. DO NOT over-tighten hardware attaching to the frame. This could cause damage to the wheelchair.


# Procedure 1 - Upholstery

The upholstery is fully adjustable and can be easily adjusted to provide the correct balance and comfort in your new wheelchair.



## 1.1: Adjusting Seat

Remove the cushion from the wheelchair. The seat is adjusted by increasing or decreasing the tension in the straps on the underside of the seat (Fig 15). To remove, release all the straps and remove.



## 1.2: Adjusting Backrest

Remove the cushion and center pad from wheelchair (Fig 14). The backrest is adjusted by increasing or decreasing the tension on the straps at the back of the backrest. To remove, release all the straps and remove.

## 1.3: Changing Cushion Cover

Remove cushion from the wheelchair. Open the hook and loop pocket at the rear of the cushion and remove or replace foam as required. (Fig16).



### **⚠ WARNING:**

The position of the footrest, camber tube and rear seat height, as well as the tautness of the back upholstery are directly related to the chair's stability. Any change to one or any combination of the four may cause the chair to decrease in stability. Use **EXTREME** caution when using a new seating position for the first time.

### **⚠ CAUTION:**

Hook and loop patches used on upholstery may cause property damage if contacted closely.


# Procedure 2 - Rear Wheels

The rear wheels are quick release and are designed to be removed and attached frequently. All rear wheels use quick release axle pins to secure the wheels to the frame, although these vary in length, they operate the same.

## 2.1: Removing/attaching rear wheels

Push the quick release button and pull the wheel outwards. To install, push and hold the button and insert the quick release axle pin into the axle until the pin clicks into position. Ensure both wheels are fully engaged prior to use.

Fig 18.




### **⚠ WARNING:**

Changing the size of your rear wheels can seriously affect the performance and stability of the wheelchair.

## 2.2: Adjusting Quick release axle pin

Only adjust the quick release axle pin if the wheel does not engage or the wheels feel loose. To adjust the length, remove the axle pin from the wheel and by using 2 x adjustable spanners/wrenches as per Figure 19, extend or shorten the pin as required.

Fig 19.



A = Lengthen    B = Shorten

### **⚠ WARNING:**

Any adjustments to your quick release pins – then YOU must ensure your wheels lock on both sides of your wheelchair.

When attaching the rear wheels, always check that they are securely mounted. Pull on the rear wheels to make sure the detent pin / RGK cap of the quick release axle are fully released before operating the wheelchair.

You should NOT be able to pull the rear wheel off.

# Procedure 3 - Tires and Inner Tubes

Tires will wear just like a car tire and the time will depend on the type of use. It is recommended that both tires are changed once the tread fades. For pneumatic tires, there is a risk that a puncture will occur and the inner tube will need to be replaced in order to keep the wheelchair operational.

## 3.1: Changing the tire

Remove the wheel from the wheelchair and release air in the tube by pressing the small pin inside the valve. If a Presta valve, then unscrew both the head and the securing ring at base and press the top until air releases. Using 2 x tire levers (supplied in the toolkit), unhook the tire completely from the wheel rim. Remove inner tube and tire from wheel rim.

To install the tire, place one side of the new tire on the wheel rim. Then insert the inner tube inside the tire and align the valve with the wheel rim hole. Ensure there are no kinks or uneven tension in the inner tube before putting the second side of the tire back on the wheel rim using the

tire levers. Be careful not to pinch the inner tube. Once tire and inner tube are on the wheel rim, replace securing ring (if applicable) then inflate tire a little. Adjust inner tube and tire so that they fit nicely on the wheel rim and inflate to the recommended tire pressure.

## 3.2: Changing the inner tube

Repeat the method above but do not remove tire completely, just remove inner tube. If problems continue, remove tire and inner tube and ensure there are no sharp edges in the wheel rim which may be causing a puncture.

# Procedure 4 - Camber Bar

The Camber Axle or Camber Tube control your wheel position and your center of gravity. It is very important to have this set up for your requirements for performance and safety. Axle tube can be fixed or adjustable. Fixed means this cannot be adjusted and for a user who is experienced and knows their balance point. Adjustable position gives the user choice and is ideal for less experienced users

## 4.1: Adjusting Axle Tube position

To adjust the axle position, remove wheels. Using a 5mm allen key from the toolkit, release the top screw on each side of the axle clamp as shown in Fig 21 (A). Once released but still intact, move the axle backward for more stability and move forward for more tip. Tighten the top screws once position has been set, and carefully use the chair as if it were the first time.

## 4.2: Replacing Axle tube

To replace the axle tube, remove wheels. Using a 5mm allen key from the toolkit, release the bottom screw completely from the axle clamp as shown in Fig 21 (B). Take note of the distance between the axle clamp and the axle tube for set up of the new axle tube. Slide new tube into the clamps and into the same position as the old axle tube. The notch in the axle insert must be at the top of the axle tube.

If the wheelchair feels slow after replacing the axle tube, rotate the axle tube a little forwards or backwards to correct the wheel alignment.

## ⚠ WARNING:

The position of the footrest, camber bar, as well as the tautness of the back upholstery, are directly related to the chair's stability. Any change to one or any combination of the three may cause the chair to decrease in stability. Use EXTREME caution when using a new seating position for the first time.




Fig 21.

A = Adjust Position  
B = Remove and replace

# Procedure 5 - Caster Wheels and Forks

The caster wheels at the front of the wheelchair will wear just like car tires and should be replaced when they visibly look worn or cracks appear.

## 5.1: Changing caster wheels

Using 2 x 4mm allen keys from the toolkit, remove 1 of the 2 screws in the center of the caster wheel, as shown in Fig 22. Remove the screw and slide out the caster axle with the other screw still attached, noting any spacers and their positions. Install the new caster wheel and slide the caster axle back through and secure with the final screw.

Fig 22.



## 5.1: Changing caster Fork

Remove the bolt at the back of the caster housing using a 4mm Allen key from the toolkit. Then use the special spanner to undo the caster fork assembly as pictured in Fig 22, while holding the frame. The entire caster assembly will unscrew giving you access to the bearings and caster fork. Use the socket to release the nut which will release the fork. When installing, screw in the housing as far as it can go and line up the hole so the securing bolt can be tightened.

### NOTE:

Caster Bearings must be fitted by an authorized dealer. Failure to do so will result in invalid warranty on the parts.

### CAUTION:

DO NOT replace the front caster forks or caster wheels with a different size other than what was originally supplied on the chair. Changing to a different size caster fork or caster wheel will have an adverse effect on the running of the chair and the end user's balance causing personal injury.

# Procedure 6 - Wheel Locks

The wheel locks are adjustable in position to suit tire changes and wear. As tires wear, the wheel lock may not hold the wheelchair as intended, so these need to be adjusted to suit the tire condition.

### NOTE:

Before adjusting or replacing any style of wheel lock ensure that the tires are inflated to the recommended pressure as indicated on the tire.

## 6.1: Adjusting Wheel locks

Using a 5mm Allen key from the toolkit, release the 2 screws in the clamp or on the underside of the plate. Once the wheel lock is loose, reposition so that there is at least 3mm of interference when the wheel lock is activated. Tighten securely and repeat on the other side.



### WARNING:

After any adjustments, repair or service and before use, make sure all attaching hardware is tightened securely – otherwise injury or damage may result. If wheel locks do not hold occupied wheelchair in place please adjust as above, or contact your dealer – otherwise injury or damages may occur. The wheel locks are designed to work when your wheelchair is static. Wheel locks should not be applied during use and tire pressure should be as indicated on the tire.

# Procedure 7 - Footrest

The footrest may need adjusting when you first use the wheelchair, if circumstances change or if fasteners come loose. It is important to periodically check the height and adjust as required.

## 7.1: Adjust footrest height


Before adjustment, take note of the height you have or require and mark if necessary. Using a 4 or 5mm allen key from the toolkit, release the 2 screws in the footrest clamps. Reposition the footrest in the correct height and tighten the screws securely.

## 7.2: Adjust the footrest angle

With an adjustable angle footrest, release the 2 x clamp screws in the center of the footplate using a 4mm allen key and 10mm spanner from the toolkit. Reposition to the correct angle and tighten the screws securely.

### **⚠ WARNING:**

The footrest assembly **MUST** be at least 1.6-in (40mm) above the ground/floor to avoid hitting protruding objects when using this wheelchair.



# Procedure 8 - Folding Backrest

The backrest can be fixed or folding and if a folding option is chosen, the backrest can fold to make the chair smaller in transport size.

## 8.1: Operating the folding backrest

Grab the pull cord and the back brace-bar together until the backrest releases and moves forward. Release the pull cord and continue to push forward until the pins click into position securing the backrest in the folded position. To release, grab the pull cord and back brace bar together and pull upright until the backrest clicks into the original position



### **⚠ WARNING:**

Your folding back **MUST** be locked securely in place before using your wheelchair.

### **⚠ CAUTION:**

This product contains moving parts. Beware of finger trapping.

# Procedure 9 - Push Handles

A range of push handles can be supplied on this wheelchair for use by an attendant and operated as per the following procedures.

## 9.1: Adjusting Height adjustable push handles

Release the over center lever and position at the desired height and then tighten the lever to secure. Repeat on the other side to get uniform height between both handles.



Height Adjustable push handles

## 9.2: Screw in push handles

Screw in push handles can be removed by unscrewing them from the backrest. To install, screw them in until tight by hand.



Screw-In push handles

## 9.3: Operating flip down push handles

Press the button underneath the handle and push the handle down at the same time, repeat on other side. Pull up the handle until it clicks into position to return to original position.



Flip-Down push handles

## **⚠ WARNING:**

DO NOT attempt to lift the wheelchair by any removable (detachable) parts i.e. Height adjustable push handles, Fold-down push handles and Screw-in push handles.

# Procedure 10 - Anti-tips

An anti tip can be used to make the chair more stable and is recommended by a health care professional.

## 10.1: Swing away anti-tip

To store the anti-tip, push down on the spring-loaded tube until it releases from the housing. Swing 180 degrees under the chair and into the locked position. To return to the operating position, repeat and return the same way. To adjust the length, press the button on the tube and lengthen or shorten the wheel tube.

## 10.2: Single sport anti-tip

To remove anti-tip, remove quick release pin and remove tube. The clamp attached to the frame is factory fitted and cannot be removed unless by an approved wheelchair technician.

### **⚠ WARNING:**

Sunrise Medical strongly recommends ordering the anti-tips as an additional safeguard for the wheelchair user.

### **⚠ WARNING:**

Anti-tippers **MUST** be fully engaged. Ensure the release button of the anti-tipper fully protrudes out of the hole in the anti-tipper socket.

Ensure both anti-tippers are adjusted to the same height.






# Procedure 11 - Sideguards

Sideguards support the user and prevent clothes from getting dirty. Although these need little maintenance, replacement is shown below.

## 11.1: Screw in sideguards

Remove the rear wheels. Using a 4mm allen key from the toolkit, release the 3 screws in the sideguard. To install, secure all screws but do not tighten until all are in position. Take care not to cross thread the screws.



## 11.2: Lift out sideguards

Remove sideguards from receiver. Using a 4mm allen key from the toolkit, release the 2 screws securing the sideguard to the lift out post. To install, replace and tighten the 2 screws. Take care not to cross thread the screws.




## 11.3: Fold flat sideguards

Using a 3mm allen key from the toolkit, remove the 4 screws from the inside of the sideguard. To install, secure all screws but do not tighten until all are in position. Take care not to cross thread the screws.



## 11.4: Flip back sideguards

Using a 4mm allen key from the toolkit, release the screws securing the sideguard to the lift out post. To install, replace and tighten the 2 screws. Take care not to cross thread the screws.



# Transportation

## **DANGER:**

There is risk of serious injury or death if this advice is ignored.

### Transportation of your wheelchair within a vehicle

A wheelchair secured in a vehicle, will not provide the equivalent of safety and security as a vehicle seating system. We recommend that the user transfers to the vehicle seating. It is recognized that this is not always practical for the user to be transferred and in these circumstances Sunrise Medical recommends the use of a crash-tested wheelchair.

## Technical Data

ITEM	Torque
Backrest Pivot Plate	10Nm [88.5 in-lbs]
Wheel Lock	10Nm [88.5 in-lbs]
Camber Bar	10Nm [88.5 in-lbs]
Footrest Adjustment Height	7Nm [62 in-lbs]
All Other Items	7Nm [62 in-lbs]

## NOTE:

RGK wheelchairs are bespoke and individually made to measure for you, therefore your wheelchair may fall below or over the mentioned sizes.

## Recycling/Disposal

Prior to the wheelchair being re-used, it must be carefully prepared. All surfaces which come into contact with the user must be cleaned down and treated with a disinfection spray.

In the following section, there is a description of the materials used on our wheelchairs.

**Aluminum:** Frame, Caster forks, Wheels, Side guards, Armrest, Footrest, Handrims, Anti-Tippers, Frame, Footplate covers

**Steel:** Frame, Quick release axle pins, Footrest, Anti-tippers, Frame Wheels, Hardware, Wheel Locks, Footplate covers

**Titanium:** Frame, Caster Forks, Quick release axle, Hardware, Side guards, Anti-Tippers, Frame, Footrest

**Plastic:** Handles, Tube stoppers, Casters, Footplate covers, Arm pads

**Packing:** Plastic bags made from polyethylene, cardboard

**Upholstery:** Woven polyester, Cotton, Leather



Disposal or recycling should be carried out by a disposal company or at a public disposal point. You could also return your wheelchair to your dealer for disposal (There may be a charge imposed).

# Manufacturers Warranty

**THIS GUARANTEE DOES NOT AFFECT YOUR LEGAL RIGHTS IN ANY WAY.**

Sunrise Medical\* provides a guarantee, as set out in the warranty conditions, for wheelchairs to its customers covering the following

**Warranty conditions**

1. Should a part or parts of the wheelchair require repair or replacement as a result of a manufacturing and/or material fault within 12 months or for frame and cross-braces within 5 years after delivery to the customer, then the affected part or parts will be repaired or replaced free of charge.
2. To enforce the warranty, please contact the local reseller where the product was purchased with the exact details of the nature of the difficulty. Should you be using the wheelchair outside the area covered by the local reseller, repairs or replacement will be carried out by another agency as designated by the manufacturer the wheelchair must be repaired by a Sunrise Medical authorized dealer.
3. For parts, which have been repaired or exchanged within the scope of this warranty, we provide a warranty in accordance with these warranty conditions for the remaining warranty period for the wheelchair in accordance with point 1).
4. For original spare parts which have been fitted at the customer's expense, these will have a 12 month guarantee, (following the fitting), in accordance with these warranty conditions.
5. Consumable parts such as pads, tires, inner tubes and similar parts are excluded from the warranty, except in the case that premature wear of the part is the direct result of a manufacturing fault.
6. Claims from this warranty shall not arise, if a repair or replacement of a wheelchair or a part is required for the following reasons:
  - The product or part has not been maintained or serviced in accordance with the manufacturer's recommendations as shown in the User Instructions and/or the Service Instructions.
  - Accessories have been used which are not specified as original accessories.
  - The wheelchair or part having been damaged by neglect, accident or improper use.
  - Changes/modifications have been made to the wheelchair or parts, which deviate from the manufacturer's specifications.
  - Repairs have been carried out, before our Customer Service has been informed of the circumstances.
7. This guarantee is subject to the law of the country in which the product was purchased from Sunrise Medical.

\* Means the Sunrise Medical authorized dealer from which the product was purchased.

---

Record your serial number here for future reference.

The serial number of your made to measure wheelchair is located on either the under seat cross brace bar (J Product Description) or camber bar (H Product Description) or inside the backrest upholstery (A Product Description)). (Fig. 2, page 4)

Sunrise Medical S.r.l.  
Via Riva, 20 – Montale  
29122 Piacenza  
Italia  
Tel.: +39 0523 573111  
Fax: +39 0523 570060  
www.SunriseMedical.it

Sunrise Medical AG Lückhalde 14  
3074 Muri/Bern  
Schweiz/Suisse/Svizzera  
Fon +41 (0)31 958 3838  
Fax +41 (0)31 958 3848  
www.SunriseMedical.ch

Sunrise Medical AS  
Delitoppen 3  
1540 Vestby  
Norway  
Telefon: +47 66 96 38 00  
Faks: +47 66 96 38 80  
post@sunrisemedical.no  
www.SunriseMedical.no

Sunrise Medical AB  
Neogatan 5  
431 53 Mölndal  
Sweden  
Tel: +46 (0)31 748 37 00  
post@sunrisemedical.se  
www.SunriseMedical.se

MEDICCO s.r.o.  
H – Park, Heršpická 1013/11d,  
625 00 Brno  
Czech Republic  
Tel.: (+420) 547 250 955  
Fax: (+420) 547 250 956  
www.medicco.cz  
info@medicco.cz  
Bezplatná linka 800 900 809

Sunrise Medical ApS  
Mårkærvej 5-9  
2630 Taastrup  
Denmark  
+45 70 22 43 49  
Email: info@sunrisemedical.dk  
www.SunriseMedical.dk

Sunrise Medical Pty. Ltd.  
11 Daniel Street, Wetherill Park,  
NSW 2164,  
Australia  
Phone: 9678 6600,  
Orders Fax: 9678 6655, Admin Fax: 9831  
2244. Australia  
www.SunriseMedical.com.au

Sunrise Medical (US) LLC  
North American Headquarters  
2842 N. Business Park Avenue  
Fresno, CA, 93727, USA  
(800) 333-4000  
(800) 300-7502  
www.SunriseMedical.com



Sunrise Medical GmbH  
Kahlbachring 2-4  
69254 Malsch/Heidelberg  
Deutschland  
Tel.: +49 (0) 7253/980-0  
Fax: +49 (0) 7253/980-222  
www.SunriseMedical.de

Sunrise Medical Ltd.  
Thorns Road  
Brierley Hill  
West Midlands  
DY5 2LD  
England  
Phone: 0845 605 66 88  
Fax: 0845 605 66 89 www.  
SunriseMedical.co.uk

Sunrise Medical S.L.  
Polígono Bakiola, 41  
48498 Arrankudiaga – Vizcaya  
España  
Tel.: +34 (0) 902142434  
Fax: +34 (0) 946481575  
www.SunriseMedical.es

Sunrise Medical Poland  
Sp. z o.o.  
ul. Elektronowa 6,  
94-103 Łódź  
Polska  
Telefon: + 48 42 275 83 88  
Fax: + 48 42 209 35 23  
E-mail: pl@SunriseMedical.de  
www.Sunrise-Medical.pl

Sunrise Medical B.V.  
Groningenhaven 18-20  
3433 PE NIEUWEGEIN  
The Netherlands  
T: +31 (0)30 – 60 82 100  
F: +31 (0)30 – 60 55 880  
E: info@SunriseMedical.nl  
www.SunriseMedical.nl

Sunrise Medical HCM B.V.  
Vossenbeemd 104  
5705 CL Helmond  
The Netherlands  
T: +31 (0)492 593 888  
E: customerservice@SunriseMedical.nl  
www.SunriseMedical.nl  
www.SunriseMedical.eu (International)

Sunrise Medical S.A.S  
ZAC de la Vrillonnerie  
17 Rue Mickaël Faraday  
37170 Chambray-Lès-Tours  
Tel : + 33 (0) 2 47 55 44 00  
Email: info@SunriseMedical.fr  
www.SunriseMedical.fr

Sunrise Medical Canada Inc.  
237 Romina Drive, Unit 3  
Concord , Ontario L4K 4V3  
Canada  
Phone: 1-800-263-3390  
Fax: 1-800-561-5834  
www.SunriseMedical.ca

RGK Wheelchairs Ltd.  
Unit 8 A/B/C Ring Road Zone 2  
Burntwood Industrial Estate  
Burntwood, Staffordshire, WS7 3JQ  
United Kingdom  
Tel: +44 (0) 1543 670 077



© 2024 Sunrise Medical (US) LLC  
256936 Rev. C